

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021721\
 Data File : PS014034.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Feb 2021 11:47
 Operator : DD\AJ
 Sample : M1395-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-BPM-03-020-ABCDMS

Manual Integrations
 APPROVED

Ankita
 2/19/2021 9:27:43 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 17 12:16:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 16 01:39:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.901	260.6E6	105.8E6	259.950m	276.694m
Target Compounds							
1) T	Dalapon	3.023	2.616	299.3E6	193.6E6	256.781	377.305 #
2) T	3,5-DICHL...	6.421	6.059	318.3E6	121.0E6	231.404	209.188
3) T	4-Nitroph...	7.039	6.519	115.8E6	56568562	227.237	230.804
5) T	DICAMBA	7.392	7.067	920.2E6	372.2E6	229.442m	228.337m
6) T	MCPD	7.556	7.157	67951666	28608733	23.304	21.879
7) T	MCPA	7.704	7.369	99654385	60473070	23.639m	31.327m#
8) T	DICHLORPROP	8.084	7.692	231.7E6	104.8E6	229.029	248.707
9) T	2,4-D	8.315	7.990	282.1E6	131.5E6	241.305m	250.714
10) T	Pentachlo...	8.621	8.414	2986.3E6	1168.1E6	203.734	213.655
11) T	2,4,5-TP ...	9.168	8.773	1418.5E6	574.8E6	243.392	254.771m
12) T	2,4,5-T	9.455	9.156	1393.2E6	531.6E6	233.151	243.626
13) T	2,4-DB	10.007	9.661	184.5E6	72890991	172.108m	247.775 #
14) T	DINOSEB	11.157	9.993	88891895	48835696	19.767m	31.909 #
15) T	Picloram	10.993	10.983	1892.8E6	772.1E6	229.925	235.811
16) T	DCPA	11.428	10.914	1726.2E6	764.4E6	231.864	262.291

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021721\
Data File : PS014034.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Feb 2021 11:47
Operator : DD\AJ
Sample : M1395-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
FK-BPM-03-020-ABCDMS

Manual Integrations
APPROVED

Ankita
2/19/2021 9:27:43 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 17 12:16:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
Quant Title : 8080.M
QLast Update : Tue Feb 16 01:39:40 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

