

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021722\
 Data File : PS018360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Feb 2022 17:45
 Operator : AJ\MA
 Sample : N1552-02MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 02:50:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020822.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 09 15:57:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.027	6.636	928.4E6	266.2E6	590.195	424.416 #
Target Compounds							
1) T	Dalapon	2.888	2.403	217.7E6	595.6E6	86.671	743.143m#
2) T	3,5-DICHL...	6.269	5.857	310.3E6	110.6E6	138.703	116.545
5) T	DICAMBA	7.208	6.788	1178.2E6	525.1E6	186.518	193.560
6) T	MCP	7.372	6.877	49597438	26736832	10.777	12.330
7) T	MCPA	7.518	7.084	96384605	68750198	14.794	22.275 #
8) T	DICHLORPROP	7.894	7.399	222.6E6	102.0E6	141.055	147.242
9) T	2,4-D	8.124	7.688	399.6E6	189.5E6	217.046	228.646m
10) T	Pentachlo...	8.416	8.098	1312.0E6	551.2E6	62.084	63.848
11) T	2,4,5-TP ...	8.966	8.460	869.7E6	396.9E6	101.788	107.149
12) T	2,4,5-T	9.251	8.841	1124.1E6	548.3E6	127.029	154.861
13) T	2,4-DB	9.833	9.348	384.3E6	345.1E6	218.349m	724.875 #
15) T	Picloram	10.771	10.642	3054.4E6	1621.7E6	267.349	323.570
16) T	DCPA	11.205	10.584	3100.7E6	1553.7E6	281.307	341.390

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021722\
Data File : PS018360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Feb 2022 17:45
Operator : AJ\MA
Sample : N1552-02MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 18 02:50:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020822.M
Quant Title : 8080.M
QLast Update : Wed Feb 09 15:57:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

