

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021924\
 Data File : PS025988.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Feb 2024 15:11
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/20/2024
 Supervised By : Ankita Jodhani 02/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 00:18:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021424.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 14 15:17:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.270	7.802	1986.2E6	615.1E6	750.483	704.708
Target Compounds							
1) T	Dalapon	2.676	2.726	2730.8E6	985.2E6	610.443m	604.166m
2) T	3,5-DICHL...	6.437	6.750	2690.5E6	878.2E6	693.045	657.021
3) T	4-Nitroph...	7.070	7.329	1078.6E6	313.0E6	680.283	649.277
5) T	DICAMBA	7.459	8.004	7906.1E6	2618.3E6	717.538	677.629
6) T	MCP	7.642	8.104	606.1E6	200.1E6	76.699m	74.198
7) T	MCPA	7.794	8.350	806.5E6	269.6E6	70.903	69.974
8) T	DICHLORPROP	8.172	8.720	2027.0E6	646.0E6	718.490	670.626
9) T	2,4-D	8.405	9.052	2105.3E6	635.6E6	712.032	670.553
10) T	Pentachlo...	8.704	9.582	29527.8E6	9839.7E6	745.103m	684.627
11) T	2,4,5-TP ...	9.289	9.957	11443.3E6	3749.3E6	734.166	689.061
12) T	2,4,5-T	9.583	10.378	11177.1E6	3447.4E6	734.943	690.463
13) T	2,4-DB	10.162	10.945	1430.0E6	376.2E6	729.737	679.792
14) T	DINOSEB	11.381	11.325	8127.6E6	2533.2E6	733.258	705.973
15) T	Picloram	11.185	12.424	12388.2E6	3379.6E6	692.350m	631.160m
16) T	DCPA	11.674	12.370	14635.2E6	4428.2E6	744.378m	687.607

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021924\
Data File : PS025988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Feb 2024 15:11
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/20/2024
Supervised By : Ankita Jodhani 02/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 00:18:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021424.M
Quant Title : 8080.M
QLast Update : Wed Feb 14 15:17:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

