

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022020\
 Data File : PS008907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2020 10:47
 Operator : DD\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 11:18:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 20 11:16:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.247	6.941	2103.5E6	962.5E6	969.509	983.166
Target Compounds							
1) T	Dalapon	3.048	2.629	2721.4E6	1332.4E6	893.966	890.498
2) T	3,5-DICHL...	6.455	6.084	2576.6E6	1318.3E6	902.337	901.292
3) T	4-Nitroph...	7.076	6.541	957.4E6	516.5E6	889.278	889.890
5) T	DICAMBA	7.440	7.111	7407.6E6	3795.7E6	907.877	906.235
6) T	MCPD	7.613	7.212	590.3E6	293.8E6	94.853	94.109
7) T	MCPA	7.763	7.429	776.5E6	405.2E6	93.596	80.984
8) T	DICHLORPROP	8.135	7.739	1781.6E6	921.7E6	919.418	921.792
9) T	2,4-D	8.363	8.033	1891.7E6	1089.1E6	904.421	912.075
10) T	Pentachlo...	8.677	8.467	25760.9E6	12526.0E6	889.658	883.826
11) T	2,4,5-TP ...	9.227	8.828	10000.1E6	4960.0E6	908.815	911.276
12) T	2,4,5-T	9.511	9.207	9756.4E6	4971.4E6	901.356	910.548
13) T	2,4-DB	10.065	9.716	1379.3E6	661.5E6	944.624	932.517
14) T	DINOSEB	11.231	10.051	7628.2E6	3569.0E6	898.840	894.523
15) T	Picloram	11.044	11.032	13139.5E6	6961.3E6	933.577	926.166
16) T	DCPA	11.495	10.983	12570.9E6	6311.6E6	920.280	906.985

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022020\
Data File : PS008907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2020 10:47
Operator : DD\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 11:18:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
Quant Title : 8080.M
QLast Update : Thu Feb 20 11:16:54 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

