

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022023\
 Data File : PS022023.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2023 12:59
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 13:24:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:05:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.298	6.723	3489.8E6	3025.7E6	1424.866	1527.935
Target Compounds							
1) T	Dalapon	2.408	2.483	4428.0E6	4527.4E6	1378.994	1379.122
2) T	3,5-DICHL...	5.678	5.923	4988.6E6	4257.7E6	1334.896	1362.153
3) T	4-Nitroph...	6.138	6.347	1938.7E6	1889.6E6	1476.057	1397.080
5) T	DICAMBA	6.434	6.882	13726.5E6	12990.5E6	1312.932	1326.861
6) T	MCPD	6.594	6.989	1146.9E6	1147.3E6	146.187	146.408
7) T	MCPA	6.718	7.201	1508.0E6	1465.1E6	141.510	141.583
8) T	DICHLORPROP	7.019	7.502	3596.7E6	3363.7E6	1344.318	1380.303
9) T	2,4-D	7.219	7.792	4143.8E6	3948.3E6	1249.604	1351.557
10) T	Pentachlo...	7.440	8.214	38714.3E6	37924.9E6	1060.474	1167.965
11) T	2,4,5-TP ...	7.967	8.575	19938.7E6	17645.4E6	1293.977	1307.215
12) T	2,4,5-T	8.226	8.949	21547.0E6	16677.5E6	1245.470	1362.298
13) T	2,4-DB	8.731	9.454	4307.6E6	2488.0E6	1537.869	1430.627
14) T	DINOSEB	9.777	9.778	17172.9E6	12918.3E6	1279.286	1355.129
15) T	Picloram	9.612	10.745	33806.8E6	26595.9E6	1345.004	1319.672
16) T	DCPA	10.045	10.703	27141.2E6	20354.4E6	1287.805	1303.055

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022023\
Data File : PS022203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2023 12:59
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 13:24:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

