

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022023\
 Data File : PS022209.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2023 17:16
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/21/2023
 Supervised By : Ankita Jodhani 02/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 00:57:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:26:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.304	6.722	1811.8E6	1504.8E6	730.950	716.945m
Target Compounds							
1) T	Dalapon	2.408	2.477	2198.7E6	2261.6E6	693.534	702.341
2) T	3,5-DICHL...	5.684	5.922	2597.3E6	2199.1E6	698.698	711.132
3) T	4-Nitroph...	6.148	6.349	936.8E6	944.1E6	736.196	706.912
5) T	DICAMBA	6.440	6.881	7324.0E6	6957.9E6	705.816	719.985
6) T	MCPD	6.593	6.980	537.8E6	561.6E6	70.478	73.566
7) T	MCPA	6.716	7.190	738.5E6	725.5E6	70.916	71.178
8) T	DICHLORPROP	7.027	7.502	1883.5E6	1726.8E6	697.423	702.383
9) T	2,4-D	7.230	7.793	2333.2E6	2052.9E6	711.628	708.940
10) T	Pentachlo...	7.448	8.210	26137.2E6	23306.9E6	725.977	719.310
11) T	2,4,5-TP ...	7.977	8.575	11011.5E6	9698.8E6	713.991	721.043
12) T	2,4,5-T	8.238	8.950	12163.8E6	8885.4E6	701.120	738.786
13) T	2,4-DB	8.745	9.455	1982.9E6	1251.2E6	731.430	753.614
14) T	DINOSEB	9.788	9.778	9409.3E6	6845.6E6	684.663	717.491
15) T	Picloram	9.628	10.748	18146.6E6	14574.1E6	736.647	743.347
16) T	DCPA	10.056	10.702	15116.1E6	11353.9E6	709.725	725.728

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022203\
Data File : PS022209.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2023 17:16
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/21/2023
Supervised By : Ankita Jodhani 02/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 00:57:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022203.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:26:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

