

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022223\
 Data File : PS022249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2023 18:49
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 20:18:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:26:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.300	6.723	1746.7E6	1459.5E6	704.704	695.321
Target Compounds							
1) T	Dalapon	2.408	2.482	1892.9E6	1963.4E6	597.084	609.733
2) T	3,5-DICHL...	5.681	5.923	2472.0E6	2115.1E6	664.989	683.988
3) T	4-Nitroph...	6.145	6.350	884.5E6	891.3E6	695.066	667.371
5) T	DICAMBA	6.435	6.881	7016.7E6	6690.3E6	676.205	692.289
6) T	MCPD	6.588	6.979	526.2E6	534.8E6	68.961	70.049
7) T	MCPA	6.710	7.190	707.1E6	699.5E6	67.903	68.631
8) T	DICHLORPROP	7.020	7.501	1784.0E6	1666.1E6	660.597	677.708
9) T	2,4-D	7.222	7.792	2206.9E6	1957.9E6	673.100	676.136
10) T	Pentachlo...	7.441	8.209	24860.1E6	22643.2E6	690.504	698.827
11) T	2,4,5-TP ...	7.968	8.574	10469.9E6	9376.5E6	678.877	697.085
12) T	2,4,5-T	8.228	8.949	11475.5E6	8607.2E6	661.449	715.650
13) T	2,4-DB	8.734	9.453	1742.1E6	1203.2E6	642.611	724.681
14) T	DINOSEB	9.775	9.775	8981.7E6	6495.9E6	653.547	680.844
15) T	Picloram	9.616	10.746	15227.0E6	12290.7E6	618.129	626.887
16) T	DCPA	10.044	10.700	14332.6E6	11020.6E6	672.939	704.424

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022223\
Data File : PS022249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2023 18:49
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 20:18:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:26:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

