

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022224\
 Data File : PS026003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2024 10:48
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 12:19:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022224.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 22 12:15:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.250	7.783	1320.2E6	378.6E6	517.686	501.005
Target Compounds							
1) T	Dalapon	2.670	2.722	2046.1E6	744.4E6	471.307	462.494
2) T	3,5-DICHL...	6.420	6.734	1807.2E6	550.1E6	479.725	465.300
3) T	4-Nitroph...	7.050	7.310	586.1E6	154.3E6	461.301	456.932
5) T	DICAMBA	7.438	7.984	5210.8E6	1622.5E6	477.635	461.197
6) T	MCPD	7.620	8.084	356.3E6	123.5E6	45.778	45.962
7) T	MCPA	7.770	8.328	488.1E6	156.1E6	46.125	45.465
8) T	DICHLORPROP	8.150	8.700	1332.2E6	401.1E6	484.683	469.930
9) T	2,4-D	8.382	9.030	1260.9E6	343.9E6	475.946	459.040
10) T	Pentachlo...	8.680	9.558	19694.6E6	6210.5E6	491.205	477.630
11) T	2,4,5-TP ...	9.263	9.934	7271.9E6	2215.2E6	484.231	468.797
12) T	2,4,5-T	9.556	10.354	6494.3E6	1719.5E6	474.891	455.244
13) T	2,4-DB	10.134	10.920	759.6E6	172.8E6	467.759	453.660
14) T	DINOSEB	11.349	11.298	5067.3E6	1443.8E6	470.632	457.366
15) T	Picloram	11.154	12.395	6609.8E6	1534.6E6	457.671	435.891
16) T	DCPA	11.642	12.342	9166.3E6	2458.3E6	486.767	468.740

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022224\
Data File : PS026003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2024 10:48
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 12:19:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022224.M
Quant Title : 8080.M
QLast Update : Thu Feb 22 12:15:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

