

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022423\
 Data File : PS022264.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2023 11:36
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 00:14:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:26:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.300	6.723	1815.3E6	1514.5E6	732.355	721.549
Target Compounds							
1) T	Dalapon	2.409	2.483	1990.0E6	2085.0E6	627.701	647.492
2) T	3,5-DICHL...	5.681	5.923	2575.9E6	2201.3E6	692.920	711.851
3) T	4-Nitroph...	6.145	6.350	923.5E6	935.0E6	725.763	700.070
5) T	DICAMBA	6.435	6.880	7303.7E6	6970.7E6	703.867	721.306
6) T	MCPD	6.588	6.979	550.2E6	559.5E6	72.106	73.289
7) T	MCPA	6.710	7.189	740.7E6	730.1E6	71.126	71.632
8) T	DICHLORPROP	7.020	7.500	1848.5E6	1728.5E6	684.473	703.098
9) T	2,4-D	7.222	7.792	2279.3E6	2045.4E6	695.176	706.366
10) T	Pentachlo...	7.441	8.208	25556.0E6	23375.9E6	709.832	721.440
11) T	2,4,5-TP ...	7.968	8.573	10786.5E6	9731.2E6	699.400	723.454
12) T	2,4,5-T	8.228	8.947	11881.9E6	8955.5E6	684.873	744.609
13) T	2,4-DB	8.733	9.452	1921.0E6	1253.4E6	708.605	754.945
14) T	DINOSEB	9.774	9.774	9284.1E6	6725.6E6	675.551	704.923
15) T	Picloram	9.615	10.745	17093.2E6	13933.5E6	693.888	710.675
16) T	DCPA	10.043	10.698	14686.6E6	11421.8E6	689.561	730.067

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022423\
Data File : PS022264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2023 11:36
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 00:14:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:26:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

