

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022520\
 Data File : PS009006.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2020 19:27
 Operator : DD\AJ
 Sample : L1603-01MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HAM-WC-S-SB-010203MS

Manual Integrations
 APPROVED

Sohil
 2/27/2020 2:36:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 01:13:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 20 11:30:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.246	6.940	708.5E6	365.9E6	324.019	379.697
Target Compounds							
1) T	Dalapon	3.047	2.629	987.7E6	726.5E6	330.138	496.872 #
2) T	3,5-DICHL...	6.454	6.083	1404.8E6	706.0E6	490.577	483.824m
5) T	DICAMBA	7.440	7.111	3880.7E6	2023.6E6	477.698	488.871
6) T	MCP	7.609	7.207	338.0E6	167.7E6	56.202m	55.334
7) T	MCPA	7.758	7.422	422.3E6	239.1E6	52.505	54.034
8) T	DICHLORPROP	8.134	7.739	1012.5E6	549.4E6	521.232	548.300
9) T	2,4-D	8.363	8.033	1230.8E6	722.4E6	593.920	609.700
10) T	Pentachlo...	8.676	8.466	11357.9E6	5828.2E6	391.370	410.224
11) T	2,4,5-TP ...	9.226	8.826	5950.8E6	4529.8E6	541.200	836.232 #
12) T	2,4,5-T	9.510	9.207	5471.4E6	2901.3E6	511.728	535.985
13) T	2,4-DB	10.064	9.715	666.1E6	377.0E6	461.599m	536.726
14) T	DINOSEB	11.230	10.051	2434.7E6	1167.0E6	287.777	295.130
15) T	Picloram	11.043	11.030	6040.3E6	3318.0E6	441.187	456.613
16) T	DCPA	11.495	10.981	4726.9E6	2589.5E6	344.003	374.241

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022520\
Data File : PS009006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2020 19:27
Operator : DD\AJ
Sample : L1603-01MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
Client Sample ID :
HAM-WC-S-SB-010203MS

Manual Integrations
APPROVED

Sohil
2/27/2020 2:36:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 01:13:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
Quant Title : 8080.M
QLast Update : Thu Feb 20 11:30:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

