

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022521\
 Data File : PS014083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2021 16:32
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 17:21:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 16 01:39:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.204	6.901	745.7E6	287.3E6	743.829	751.214
Target Compounds							
1) T	Dalapon	3.027	2.620	766.2E6	344.7E6	657.384	671.636
2) T	3,5-DICHL...	6.422	6.058	952.3E6	408.3E6	692.420	706.076
3) T	4-Nitroph...	7.039	6.517	337.6E6	166.4E6	662.280	679.007
5) T	DICAMBA	7.393	7.067	2852.3E6	1183.1E6	711.225	725.896
6) T	MCPD	7.560	7.160	206.5E6	95701440	70.815	73.190
7) T	MCPA	7.709	7.375	298.6E6	136.5E6	70.825	70.726
8) T	DICHLORPROP	8.084	7.691	706.9E6	302.4E6	698.910	717.360
9) T	2,4-D	8.316	7.989	820.1E6	371.7E6	701.545	708.659
10) T	Pentachlo...	8.622	8.412	10414.5E6	3989.3E6	710.508	729.679
11) T	2,4,5-TP ...	9.170	8.774	4130.7E6	1655.9E6	708.739	734.001
12) T	2,4,5-T	9.458	9.155	4235.3E6	1585.2E6	708.766	726.513
13) T	2,4-DB	10.012	9.662	712.6E6	205.7E6	664.885	699.109
14) T	DINOSEB	11.161	9.989	3262.8E6	1150.0E6	725.555	751.409
15) T	Picloram	11.000	10.986	5192.4E6	2169.1E6	630.728	662.500
16) T	DCPA	11.427	10.915	5270.0E6	2147.2E6	707.883	736.768

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022521\
Data File : PS014083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2021 16:32
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 17:21:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
Quant Title : 8080.M
QLast Update : Tue Feb 16 01:39:40 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

