

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022522\
 Data File : PS018445.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2022 22:46
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 23:04:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022122.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 22 00:27:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.024	6.637	1271.5E6	526.1E6	764.090	789.469
Target Compounds							
1) T	Dalapon	2.885	2.404	1796.4E6	575.0E6	662.254	672.904
2) T	3,5-DICHL...	6.269	5.856	1651.0E6	737.3E6	694.983	723.245
3) T	4-Nitroph...	6.859	6.278	593.4E6	312.1E6	689.287	714.341
5) T	DICAMBA	7.207	6.789	4792.6E6	2113.5E6	729.001	752.663
6) T	MCPD	7.376	6.884	366.1E6	172.2E6	75.627	75.692
7) T	MCPA	7.523	7.093	519.6E6	247.9E6	73.890	74.002
8) T	DICHLORPROP	7.891	7.399	1212.6E6	549.5E6	717.006	746.370
9) T	2,4-D	8.122	7.690	1452.4E6	694.3E6	711.971	741.716
10) T	Pentachlo...	8.414	8.096	16777.1E6	6885.4E6	708.531	722.410
11) T	2,4,5-TP ...	8.963	8.459	6967.6E6	3135.4E6	731.825	759.448
12) T	2,4,5-T	9.246	8.837	7303.5E6	2989.1E6	726.047	759.378
13) T	2,4-DB	9.796	9.339	1246.9E6	419.0E6	717.531	775.579
14) T	DINOSEB	10.932	9.657	5488.6E6	2146.2E6	706.459	738.286
15) T	Picloram	10.768	10.641	8670.1E6	3988.2E6	682.359	689.643
16) T	DCPA	11.202	10.581	8763.8E6	3727.8E6	728.788	750.759

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022522\
Data File : PS018445.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2022 22:46
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 23:04:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022122.M
Quant Title : 8080.M
QLast Update : Tue Feb 22 00:27:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

