

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022522\
 Data File : PS018449.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Feb 2022 00:20
 Operator : AJ\MA
 Sample : WATER IDOC 04
 Misc : FOR ABDUL MIRZA
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WATER IDOC 04

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/28/2022
 Supervised By :Ankita Jodhani 02/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 01:07:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022122.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 22 00:27:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.026	6.639	781.9E6	310.7E6	469.897	466.226m
Target Compounds							
1) T	Dalapon	2.888	2.405	1200.6E6	382.5E6	442.627	447.668
2) T	3,5-DICHL...	6.271	5.859	1036.2E6	453.2E6	436.179	444.570
3) T	4-Nitroph...	6.869	6.287	352.9E6	202.4E6	409.977	463.286
5) T	DICAMBA	7.207	6.790	3077.5E6	1356.0E6	468.117	482.893
6) T	MCPD	7.373	6.880	223.3E6	107.6E6	46.128	47.300
7) T	MCPA	7.519	7.088	325.7E6	154.9E6	46.319	46.244
8) T	DICHLORPROP	7.893	7.401	780.4E6	346.1E6	461.409	469.996
9) T	2,4-D	8.128	7.697	971.9E6	443.1E6	476.447	473.356
10) T	Pentachlo...	8.414	8.098	11285.1E6	4724.2E6	476.590	495.658
11) T	2,4,5-TP ...	8.965	8.461	4530.2E6	2050.7E6	475.820	496.703
12) T	2,4,5-T	9.249	8.840	4841.6E6	1895.6E6	481.313	481.565
13) T	2,4-DB	9.800	9.343	798.9E6	268.1E6	459.706	496.305
14) T	DINOSEB	10.933	9.658	3633.1E6	1359.3E6	467.631	467.609
15) T	Picloram	10.774	10.647	5997.4E6	2901.1E6	472.014	501.655
16) T	DCPA	11.202	10.582	5742.1E6	2439.5E6	477.512	491.306

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022522\
Data File : PS018449.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2022 00:20
Operator : AJ\MA
Sample : WATER IDOC 04
Misc : FOR ABDUL MIRZA
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WATER IDOC 04

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/28/2022
Supervised By : Ankita Jodhani 02/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 01:07:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022122.M
Quant Title : 8080.M
QLast Update : Tue Feb 22 00:27:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

