

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022719\
 Data File : PS003362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2019 17:24
 Operator : EI\SJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Manual Integrations
 APPROVED

Sohil
 3/1/2019 11:33:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 02:45:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 02:44:14 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.276	10.775	225.1E6	240.7E6	530.668	538.011
Target Compounds							
1) T	Dalapon	2.836	2.309	302.2E6	294.7E6	483.909	485.931
2) T	3,5-DICHL...	9.703	8.935	316.7E6	292.9E6	500.825	496.321
3) T	4-Nitroph...	10.886	9.931	101.3E6	104.4E6	457.850	478.147
5) T	DICAMBA	11.576	11.052	842.8E6	800.7E6	493.767	495.466
6) T	MCPD	11.914	11.281	62517168	56388317	48.458m	48.534m
7) T	MCPA	12.164	11.656	93173341	80700935	47.968m	47.541m
8) T	DICHLORPROP	12.769	12.179	208.5E6	197.8E6	506.014	502.480
9) T	2,4-D	13.138	12.654	217.6E6	250.2E6	475.538	494.707
10) T	Pentachlo...	13.492	13.184	3501.8E6	3275.9E6	508.859	511.970
11) T	2,4,5-TP ...	14.371	13.812	1305.6E6	1242.4E6	498.288	503.035
12) T	2,4,5-T	14.777	14.377	1288.9E6	1002.5E6	498.043	518.637
13) T	2,4-DB	15.527	15.030	109.8E6	121.9E6	462.504	484.506
14) T	DINOSEB	16.981	15.449	1115.9E6	853.9E6	509.673	496.157
15) T	Picloram	16.771	16.686	1162.9E6	1194.6E6	460.056	475.692
16) T	DCPA	17.327	16.624	1688.0E6	1539.9E6	531.331	510.529

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022719\
Data File : PS003362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2019 17:24
Operator : EI\SJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDICC500

Manual Integrations
APPROVED

Sohil
3/1/2019 11:33:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 02:45:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 02:44:14 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

