

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022719\
 Data File : PS003366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2019 19:29
 Operator : EI\SJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS022719

Manual Integrations
 APPROVED

Sohil
 3/1/2019 11:34:04 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 03:02:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.308	10.808	322.8E6	339.9E6	735.358	723.885
Target Compounds							
1) T	Dalapon	2.871	2.343	457.5E6	416.0E6	703.893	656.996
2) T	3,5-DICHL...	9.735	8.969	455.3E6	416.2E6	683.029	674.053
3) T	4-Nitroph...	10.907	9.961	151.9E6	148.5E6	722.076	660.335
5) T	DICAMBA	11.609	11.087	1218.4E6	1147.4E6	691.174	686.789
6) T	MCPD	11.951	11.318	93017679	83151950	70.849m	70.443m
7) T	MCPA	12.202	11.694	137.9E6	118.9E6	69.520m	68.953m
8) T	DICHLORPROP	12.802	12.212	297.3E6	281.1E6	687.250	680.329
9) T	2,4-D	13.167	12.687	328.7E6	360.7E6	708.596	687.154
10) T	Pentachlo...	13.526	13.218	4976.2E6	4609.0E6	704.401	699.855
11) T	2,4,5-TP ...	14.405	13.845	1893.8E6	1781.1E6	704.955	699.467
12) T	2,4,5-T	14.808	14.409	1856.4E6	1277.6E6	702.732	633.500m
13) T	2,4-DB	15.557	15.062	173.1E6	188.0E6	730.969	736.325
14) T	DINOSEB	17.017	15.484	1566.8E6	1229.8E6	697.816	688.992
15) T	Picloram	16.800	16.719	1854.7E6	1880.4E6	753.571	740.876m
16) T	DCPA	17.363	16.659	2315.2E6	2203.0E6	702.587	708.280

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022719\
Data File : PS003366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2019 19:29
Operator : EI\SJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
ICVPS022719

Manual Integrations
APPROVED

Sohil
3/1/2019 11:34:04 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 03:02:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

