

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022720\
 Data File : PS009026.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2020 16:59
 Operator : DD\AJ
 Sample : L1370-10MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 OR-03-022520MS

Manual Integrations
 APPROVED

Sohil
 2/28/2020 11:00:19 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 00:34:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 20 11:30:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.247	6.940	745.3E6	376.0E6	340.835	390.246
Target Compounds							
1) T	Dalapon	3.047	2.630	1175.6E6	576.0E6	392.933	393.910
2) T	3,5-DICHL...	6.454	6.083	1407.5E6	714.3E6	491.517	489.497m
5) T	DICAMBA	7.440	7.110	4076.4E6	2116.8E6	501.783	511.397
6) T	MCPD	7.609	7.207	338.7E6	179.4E6	56.322m	59.186
7) T	MCPA	7.758	7.422	472.9E6	254.6E6	58.787	57.557
8) T	DICHLORPROP	8.134	7.738	1061.0E6	634.0E6	546.205m	632.760
9) T	2,4-D	8.363	8.032	1248.7E6	751.3E6	602.554	634.121
10) T	Pentachlo...	8.677	8.466	11521.5E6	5873.5E6	397.006	413.409
11) T	2,4,5-TP ...	9.226	8.826	6278.3E6	4753.4E6	570.984	877.511 #
12) T	2,4,5-T	9.511	9.206	5715.9E6	3023.5E6	534.594	558.553
13) T	2,4-DB	10.064	9.715	703.2E6	399.8E6	487.278	569.208
14) T	DINOSEB	11.231	10.051	2508.3E6	1208.6E6	296.482	305.651
15) T	Picloram	11.043	11.031	4454.5E6	2425.5E6	325.358	333.791
16) T	DCPA	11.496	10.981	5078.4E6	2767.7E6	369.586	399.993

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022720\
Data File : PS009026.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2020 16:59
Operator : DD\AJ
Sample : L1370-10MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
OR-03-022520MS

Manual Integrations
APPROVED

Sohil
2/28/2020 11:00:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 00:34:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
Quant Title : 8080.M
QLast Update : Thu Feb 20 11:30:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

