

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022724\
 Data File : PS026032.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2024 09:24
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 03:14:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022224.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 22 12:24:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.243	7.775	1763.3E6	513.7E6	678.992	675.626
Target Compounds							
1) T	Dalapon	2.666	2.719	2706.2E6	983.7E6	605.990	599.271
2) T	3,5-DICHL...	6.414	6.727	2391.9E6	734.0E6	626.433	618.550
3) T	4-Nitroph...	7.043	7.303	876.2E6	232.6E6	679.099	679.126
5) T	DICAMBA	7.431	7.976	7040.6E6	2212.0E6	642.755	632.045
6) T	MCPD	7.613	8.077	460.8E6	169.9E6	60.160	63.845
7) T	MCPA	7.764	8.321	684.3E6	220.8E6	64.532	64.221
8) T	DICHLORPROP	8.142	8.691	1800.9E6	548.7E6	643.775	641.891
9) T	2,4-D	8.373	9.021	1833.6E6	513.0E6	683.920	684.681
10) T	Pentachlo...	8.672	9.549	26458.5E6	8367.3E6	674.868	648.852
11) T	2,4,5-TP ...	9.254	9.923	10134.6E6	3136.2E6	674.651	669.427
12) T	2,4,5-T	9.547	10.343	9770.1E6	2714.0E6	716.737	726.871
13) T	2,4-DB	10.123	10.909	1205.3E6	288.2E6	738.532	751.835
14) T	DINOSEB	11.338	11.287	7308.3E6	2131.1E6	678.449	679.082
15) T	Picloram	11.143	12.381	11249.8E6	2786.1E6	781.588	795.177
16) T	DCPA	11.631	12.329	12878.8E6	3594.1E6	686.089	692.760

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022724\
Data File : PS026032.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2024 09:24
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 03:14:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022224.M
Quant Title : 8080.M
QLast Update : Thu Feb 22 12:24:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

