

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022819\
 Data File : PS003374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2019 16:05
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/1/2019 5:40:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 03:42:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.314	10.814	331.5E6	346.0E6	755.091	736.938
Target Compounds							
1) T	Dalapon	2.876	2.349	447.8E6	425.7E6	688.944	672.364
2) T	3,5-DICHL...	9.740	8.974	472.4E6	424.1E6	708.663	686.822
3) T	4-Nitroph...	10.915	9.967	156.5E6	149.7E6	743.561	665.825
5) T	DICAMBA	11.615	11.093	1249.4E6	1168.4E6	708.803	699.368
6) T	MCPD	11.957	11.324	95599455	85020322	72.815m	72.026m
7) T	MCPA	12.207	11.700	142.8E6	122.7E6	71.997m	71.126m
8) T	DICHLORPROP	12.808	12.219	299.7E6	286.7E6	692.662	693.824
9) T	2,4-D	13.174	12.693	333.8E6	366.0E6	719.607	697.240
10) T	Pentachlo...	13.532	13.224	5093.6E6	4691.5E6	721.020	712.376
11) T	2,4,5-TP ...	14.411	13.852	1935.9E6	1809.9E6	720.629	710.740
12) T	2,4,5-T	14.813	14.416	1883.2E6	1388.0E6	712.873	688.225m
13) T	2,4-DB	15.562	15.069	176.0E6	186.3E6	743.337	729.548
14) T	DINOSEB	17.022	15.489	1594.8E6	1251.1E6	710.302	700.948
15) T	Picloram	16.806	16.723	1840.4E6	1766.0E6	747.755	695.779m
16) T	DCPA	17.367	16.664	2348.8E6	2249.4E6	712.798m	723.188

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022819\
Data File : PS003374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2019 16:05
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
3/1/2019 5:40:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 03:42:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

