

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022819\
 Data File : PS003376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2019 17:10
 Operator : EI\SJ
 Sample : PB117459BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB117459BS

Manual Integrations
 APPROVED

Sohil

3/1/2019 5:40:40 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 03:43:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.308	10.809	330.9E6	339.4E6	753.803	722.749
Target Compounds							
1) T	Dalapon	2.873	2.345	399.9E6	341.7E6	615.264	539.632
2) T	3,5-DICHL...	9.734	8.969	457.3E6	422.2E6	686.013	683.770
3) T	4-Nitroph...	10.905	9.961	239.3E6	237.3E6	1137.219	1055.478
5) T	DICAMBA	11.611	11.088	1139.9E6	1081.6E6	646.696	647.400
6) T	MCPD	11.951	11.319	173.8E6	129.2E6	132.407	109.431m
7) T	MCPA	12.201	11.693	221.5E6	162.3E6	111.646	94.070m
8) T	DICHLORPROP	12.802	12.214	303.0E6	303.1E6	700.233	733.525
9) T	2,4-D	13.167	12.686	337.2E6	374.1E6	726.978	712.794m
10) T	Pentachlo...	13.527	13.219	4638.7E6	4331.4E6	656.634	657.692
11) T	2,4,5-TP ...	14.405	13.845	1716.2E6	1652.0E6	638.854	648.735
12) T	2,4,5-T	14.807	14.410	1680.0E6	1196.3E6	635.950	593.198m
13) T	2,4-DB	15.554	15.063	237.0E6	258.3E6	1000.991m	1011.407m
14) T	DINOSEB	17.016	15.484	1270.9E6	1058.4E6	566.029	592.971
15) T	Picloram	16.800	16.718	1482.1E6	1314.4E6	602.158	517.846m
16) T	DCPA	17.360	16.660	1199.1E6	1366.0E6	363.906m	439.171

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022819\
Data File : PS003376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2019 17:10
Operator : EI\SJ
Sample : PB117459BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB117459BS

Manual Integrations
APPROVED

Sohil

3/1/2019 5:40:40 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 03:43:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

