

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030121\
 Data File : PS014103.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2021 14:11
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 14:26:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 16 01:39:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.205	6.902	791.4E6	304.3E6	789.369	795.535
Target Compounds							
1) T	Dalapon	3.028	2.620	771.1E6	350.6E6	661.575	683.244
2) T	3,5-DICHL...	6.422	6.059	1003.2E6	431.2E6	729.416	745.702
3) T	4-Nitroph...	7.040	6.518	358.4E6	175.7E6	703.080	716.737
5) T	DICAMBA	7.394	7.068	3014.5E6	1256.3E6	751.679	770.840
6) T	MCPD	7.561	7.161	227.0E6	102.1E6	77.859	78.062
7) T	MCPA	7.711	7.376	317.2E6	144.9E6	75.232	75.080
8) T	DICHLORPROP	8.085	7.692	750.7E6	320.9E6	742.160	761.316
9) T	2,4-D	8.318	7.991	862.1E6	395.4E6	737.431	753.745
10) T	Pentachlo...	8.623	8.413	10934.5E6	4213.7E6	745.985	770.723
11) T	2,4,5-TP ...	9.172	8.775	4364.1E6	1760.2E6	748.789	780.228
12) T	2,4,5-T	9.460	9.156	4439.1E6	1663.2E6	742.877	762.262
13) T	2,4-DB	10.015	9.663	731.8E6	215.7E6	682.789	733.321
14) T	DINOSEB	11.164	9.989	3441.2E6	1206.2E6	765.221	788.142
15) T	Picloram	11.005	10.991	5151.1E6	2170.0E6	625.707	662.788
16) T	DCPA	11.429	10.916	5580.6E6	2278.1E6	749.610	781.678

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030121\
Data File : PS014103.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2021 14:11
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 14:26:25 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
Quant Title : 8080.M
QLast Update : Tue Feb 16 01:39:40 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

