

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030221\
 Data File : PS014104.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2021 08:39
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 18:04:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 16 01:39:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
15) T	Picloram	10.994	0.000	10723424	0	1.303 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030221\
Data File : PS014104.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2021 08:39
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 18:04:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
Quant Title : 8080.M
QLast Update : Tue Feb 16 01:39:40 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T.: 7.204 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
HEXANE




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T.: 6.901 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.994 min
Delta R.T.: -0.001 min
Response: 10723424
Conc: 1.30 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T.: 10.984 min
Response: 0
Conc: N.D.