

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030223\
 Data File : PS022300.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2023 16:50
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 21:44:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:26:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.298	6.721	1744.8E6	1479.3E6	703.924	704.795
Target Compounds							
1) T	Dalapon	2.407	2.482	2054.6E6	2132.5E6	648.090	662.241
2) T	3,5-DICHL...	5.679	5.921	2477.9E6	2129.1E6	666.574	688.495
3) T	4-Nitroph...	6.142	6.348	902.3E6	911.5E6	709.042	682.530
5) T	DICAMBA	6.434	6.878	7034.8E6	6744.5E6	677.951	697.906
6) T	MCPD	6.587	6.978	530.1E6	537.8E6	69.474	70.442
7) T	MCPA	6.709	7.188	712.3E6	705.6E6	68.393	69.229
8) T	DICHLORPROP	7.018	7.499	1795.9E6	1685.1E6	665.003	685.456
9) T	2,4-D	7.219	7.789	2073.1E6	2000.2E6	632.300	690.737
10) T	Pentachlo...	7.440	8.207	24852.7E6	22745.9E6	690.298	701.995
11) T	2,4,5-TP ...	7.966	8.571	10461.6E6	9411.5E6	678.337	699.682
12) T	2,4,5-T	8.225	8.946	11395.1E6	8774.2E6	656.814	729.536
13) T	2,4-DB	8.730	9.450	2007.9E6	1217.0E6	740.654	732.996
14) T	DINOSEB	9.773	9.774	8898.2E6	6623.4E6	647.473	694.203
15) T	Picloram	9.611	10.743	17247.2E6	13825.3E6	700.137	705.155
16) T	DCPA	10.041	10.698	14366.1E6	11033.9E6	674.512	705.273

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022300.D
Data File : PS022300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2023 16:50
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 21:44:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:26:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

