

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030321\
 Data File : PS014120.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2021 14:28
 Operator : DD\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 14:55:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030321.M
 Quant Title : 8080.M
 QLast Update : Wed Mar 03 14:53:09 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.204	6.900	1023.5E6	387.2E6	1041.982	1044.854
Target Compounds							
1) T	Dalapon	3.026	2.619	1126.7E6	484.2E6	962.349	950.071
2) T	3,5-DICHL...	6.421	6.057	1311.7E6	544.6E6	962.582	965.426
3) T	4-Nitroph...	7.037	6.514	492.4E6	231.1E6	962.648	949.996
5) T	DICAMBA	7.393	7.067	3818.4E6	1526.9E6	971.982	965.567
6) T	MCPD	7.562	7.163	292.3E6	126.2E6	100.077	98.951
7) T	MCPA	7.712	7.378	418.4E6	183.7E6	98.816	97.394
8) T	DICHLORPROP	8.084	7.691	933.5E6	385.0E6	979.321	974.171
9) T	2,4-D	8.315	7.988	1066.5E6	466.6E6	984.930	979.692
10) T	Pentachlo...	8.622	8.413	13830.5E6	5069.5E6	948.592	935.750
11) T	2,4,5-TP ...	9.170	8.774	5535.0E6	2126.6E6	975.166	970.207
12) T	2,4,5-T	9.456	9.154	5825.2E6	2141.8E6	976.105	980.060
13) T	2,4-DB	10.008	9.659	1112.5E6	287.6E6	1038.635	1001.641
14) T	DINOSEB	11.163	9.988	4413.6E6	1532.5E6	968.450	981.978
15) T	Picloram	10.993	10.981	8662.6E6	3420.8E6	1035.124	1017.385
16) T	DCPA	11.428	10.915	8251.5E6	3212.1E6	978.547	969.248

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030321\
Data File : PS014120.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2021 14:28
Operator : DD\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 14:55:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030321.M
Quant Title : 8080.M
QLast Update : Wed Mar 03 14:53:09 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

