

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030419\
 Data File : PS003404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2019 10:10
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/5/2019 2:10:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 00:01:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.279	10.778	335.4E6	355.0E6	764.062	755.941
Target Compounds							
1) T	Dalapon	2.840	2.312	460.8E6	438.4E6	709.005	692.321
2) T	3,5-DICHL...	9.707	8.939	484.3E6	432.9E6	726.628	701.107
3) T	4-Nitroph...	10.880	9.932	148.3E6	154.5E6	705.043	687.086
5) T	DICAMBA	11.581	11.057	1286.6E6	1197.1E6	729.866	716.550
6) T	MCPD	11.923	11.289	97201360	86583782	74.035m	73.350m
7) T	MCPA	12.173	11.664	142.2E6	123.8E6	71.695	71.777
8) T	DICHLORPROP	12.773	12.183	301.8E6	293.9E6	697.493m	711.189
9) T	2,4-D	13.138	12.657	322.8E6	369.8E6	695.960	704.487
10) T	Pentachlo...	13.497	13.188	5163.8E6	4730.3E6	730.959	718.272
11) T	2,4,5-TP ...	14.376	13.815	1984.4E6	1843.0E6	738.692	723.756
12) T	2,4,5-T	14.778	14.379	1920.2E6	1386.4E6	726.871	687.432
13) T	2,4-DB	15.527	15.032	172.6E6	167.3E6	728.876	655.280m
14) T	DINOSEB	16.987	15.453	1633.3E6	1278.1E6	727.445	716.069
15) T	Picloram	16.771	16.688	1809.0E6	1765.5E6	735.012	695.575
16) T	DCPA	17.332	16.628	2391.0E6	2294.1E6	725.608m	737.559

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030419\
Data File : PS003404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2019 10:10
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
Client Sampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
3/5/2019 2:10:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 00:01:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

