

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030421\
 Data File : PS014125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2021 10:57
 Operator : DD\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 14:36:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030421.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 04 14:36:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.205	6.903	200.1E6	76284338	200.061	203.383
Target Compounds							
1) T	Dalapon	3.026	2.619	226.4E6	104.7E6	194.256	205.370
2) T	3,5-DICHL...	6.423	6.061	255.5E6	105.1E6	187.622	184.935
3) T	4-Nitroph...	7.048	6.529	84541691	40115143	166.489	165.895
5) T	DICAMBA	7.393	7.067	742.3E6	305.7E6	186.323	189.541
6) T	MCPD	7.556	7.155	48301338	23913817	16.397	18.263
7) T	MCPA	7.704	7.370	74714483	37376414	17.295	19.311
8) T	DICHLORPROP	8.084	7.692	185.1E6	78417278	190.850	193.852
9) T	2,4-D	8.322	7.998	182.8E6	91122381	166.720	187.637
10) T	Pentachlo...	8.622	8.413	3008.9E6	1156.5E6	204.825	209.855
11) T	2,4,5-TP ...	9.172	8.776	1110.8E6	435.7E6	193.137	194.081
12) T	2,4,5-T	9.465	9.164	1166.0E6	402.3E6	192.151	180.880
13) T	2,4-DB	10.019	9.671	189.0E6	53032986	172.609	181.448
14) T	DINOSEB	11.162	9.990	918.3E6	303.0E6	198.222	189.925
15) T	Picloram	11.015	11.002	1339.5E6	582.2E6	157.741	168.436
16) T	DCPA	11.428	10.915	1729.8E6	695.9E6	202.882	205.410

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030421\
Data File : PS014125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2021 10:57
Operator : DD\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 14:36:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030421.M
Quant Title : 8080.M
QLast Update : Thu Mar 04 14:36:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

