

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030421\
 Data File : PS014129.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2021 12:30
 Operator : DD\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 14:37:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030421.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 04 14:36:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.900	1568.0E6	594.6E6	1567.836	1585.389
Target Compounds							
1) T	Dalapon	3.027	2.620	1682.1E6	723.9E6	1443.048	1419.538
2) T	3,5-DICHL...	6.420	6.056	1956.5E6	820.6E6	1436.616	1444.016
3) T	4-Nitroph...	7.035	6.512	758.0E6	348.5E6	1492.789	1441.362
5) T	DICAMBA	7.393	7.067	5714.0E6	2280.5E6	1434.279	1414.165
6) T	MCPD	7.565	7.167	470.1E6	197.9E6	159.583	151.148
7) T	MCPA	7.715	7.383	658.0E6	286.1E6	152.314	147.833
8) T	DICHLORPROP	8.083	7.691	1418.3E6	590.3E6	1462.342	1459.243
9) T	2,4-D	8.314	7.987	1601.6E6	713.9E6	1460.945	1470.006
10) T	Pentachlo...	8.622	8.413	19490.2E6	7116.4E6	1326.743	1291.292
11) T	2,4,5-TP ...	9.170	8.773	8128.9E6	3127.4E6	1413.333	1393.112
12) T	2,4,5-T	9.455	9.153	8551.8E6	3169.0E6	1409.323	1424.819
13) T	2,4-DB	10.006	9.659	1746.1E6	444.9E6	1594.301	1522.139
14) T	DINOSEB	11.161	9.988	6493.3E6	2274.1E6	1401.687	1425.609
15) T	Picloram	10.990	10.978	13077.2E6	5087.4E6	1539.971	1471.951
16) T	DCPA	11.428	10.915	11875.8E6	4636.9E6	1392.834	1368.608

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030421\
Data File : PS014129.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2021 12:30
Operator : DD\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 14:37:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030421.M
Quant Title : 8080.M
QLast Update : Thu Mar 04 14:36:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

