

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030421\
 Data File : PS014130.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2021 14:20
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS030421

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 15:27:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030421.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 04 14:50:18 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.205	6.896	755.8E6	283.5E6	743.383	738.367
Target Compounds							
1) T	Dalapon	3.024	2.612	773.1E6	324.7E6	639.963	608.423
2) T	3,5-DICHL...	6.422	6.053	952.7E6	400.4E6	687.659	694.258
3) T	4-Nitroph...	7.038	6.510	363.1E6	171.8E6	707.001	708.415
5) T	DICAMBA	7.394	7.063	2807.4E6	1141.1E6	698.189	700.536
6) T	MCPD	7.562	7.156	209.4E6	91243078	70.434	68.466
7) T	MCPA	7.712	7.372	300.6E6	132.5E6	68.681	66.468
8) T	DICHLORPROP	8.086	7.687	684.1E6	286.2E6	692.088	691.673
9) T	2,4-D	8.318	7.985	780.9E6	344.7E6	716.919	698.607
10) T	Pentachlo...	8.625	8.409	10460.8E6	3926.2E6	705.104	705.921
11) T	2,4,5-TP ...	9.173	8.769	4108.7E6	1598.8E6	707.492	707.745
12) T	2,4,5-T	9.460	9.151	4294.5E6	1605.0E6	702.457	723.902
13) T	2,4-DB	10.012	9.657	835.3E6	214.2E6	750.305	722.446
14) T	DINOSEB	11.166	9.985	3219.6E6	1130.0E6	682.259	699.933
15) T	Picloram	10.998	10.979	6507.9E6	2576.9E6	770.772	750.207
16) T	DCPA	11.433	10.912	6174.6E6	2456.0E6	714.753	717.094

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030421\
Data File : PS014130.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2021 14:20
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS030421

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 15:27:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030421.M
Quant Title : 8080.M
QLast Update : Thu Mar 04 14:50:18 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

