

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030521\
 Data File : PS014152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2021 14:36
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 04:26:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030521.M
 Quant Title : 8080.M
 QLast Update : Sat Mar 06 03:17:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.901	782.3E6	294.5E6	776.846	774.217
Target Compounds							
1) T	Dalapon	3.027	2.620	820.2E6	363.7E6	698.668	689.980
2) T	3,5-DICHL...	6.421	6.057	995.4E6	416.4E6	719.772	723.381
3) T	4-Nitroph...	7.038	6.516	362.0E6	171.6E6	696.913	699.735
5) T	DICAMBA	7.392	7.067	2947.1E6	1192.7E6	737.677	737.721
6) T	MCPD	7.560	7.161	220.1E6	97194267	74.603	73.328
7) T	MCPA	7.709	7.376	318.3E6	141.7E6	73.431	71.979
8) T	DICHLORPROP	8.083	7.692	713.2E6	297.8E6	730.423	727.828
9) T	2,4-D	8.316	7.990	801.7E6	355.0E6	741.481	735.425
10) T	Pentachlo...	8.621	8.413	10907.2E6	4073.4E6	743.239	742.182
11) T	2,4,5-TP ...	9.170	8.774	4246.1E6	1655.1E6	743.585	745.699
12) T	2,4,5-T	9.457	9.155	4454.6E6	1635.8E6	743.166	755.544
13) T	2,4-DB	10.011	9.662	779.7E6	214.5E6	737.219	748.587
14) T	DINOSEB	11.162	9.989	3352.8E6	1158.7E6	724.873	736.106
15) T	Picloram	10.998	10.984	6098.7E6	2472.1E6	764.070	748.836
16) T	DCPA	11.428	10.915	6324.9E6	2507.1E6	739.309	746.881

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030521\
Data File : PS014152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2021 14:36
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 04:26:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030521.M
Quant Title : 8080.M
QLast Update : Sat Mar 06 03:17:26 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

