

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030620\
 Data File : PS009184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2020 04:44
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 05:37:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 20 11:30:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.245	6.937	1686.3E6	800.9E6	771.192	831.143
Target Compounds							
1) T	Dalapon	3.049	2.632	1999.6E6	977.3E6	668.369	668.360
2) T	3,5-DICHL...	6.453	6.081	2095.9E6	1101.0E6	731.911	754.477
3) T	4-Nitroph...	7.074	6.538	778.1E6	442.8E6	727.297	766.095
5) T	DICAMBA	7.438	7.108	6105.1E6	3207.1E6	751.504	774.784
6) T	MCPD	7.609	7.207	454.4E6	247.1E6	75.551	81.514
7) T	MCPA	7.759	7.423	636.1E6	341.5E6	79.076	77.196
8) T	DICHLORPROP	8.132	7.735	1445.2E6	769.1E6	743.968	767.602
9) T	2,4-D	8.361	8.029	1496.4E6	907.1E6	722.078	765.601
10) T	Pentachlo...	8.674	8.462	20873.3E6	10703.8E6	719.250	753.397
11) T	2,4,5-TP ...	9.224	8.824	8061.2E6	4201.7E6	733.127	775.658
12) T	2,4,5-T	9.509	9.203	7674.2E6	4123.2E6	717.742	761.702
13) T	2,4-DB	10.061	9.711	1033.9E6	538.3E6	716.476	766.360
14) T	DINOSEB	11.228	10.047	6135.3E6	3034.9E6	725.186	767.538
15) T	Picloram	11.041	11.026	8117.2E6	4693.6E6	592.888	645.927
16) T	DCPA	11.492	10.976	9982.6E6	5383.7E6	726.493	778.067

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030620\
Data File : PS009184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2020 04:44
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 05:37:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022020.M
Quant Title : 8080.M
QLast Update : Thu Feb 20 11:30:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

