

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030624\
 Data File : PS026134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2024 21:27
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 03:09:05 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030624.M
 Quant Title : 8080.M
 QLast Update : Wed Mar 06 14:00:09 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.215	7.744	2068.1E6	606.4E6	764.517	782.119
Target Compounds							
1) T	Dalapon	2.653	2.701	2983.1E6	1113.3E6	690.277	695.084
2) T	3,5-DICHL...	6.390	6.702	2799.9E6	859.0E6	718.814	726.668
3) T	4-Nitroph...	7.014	7.273	1017.4E6	276.6E6	757.015	763.138
5) T	DICAMBA	7.402	7.943	8357.5E6	2630.1E6	735.941	744.841
6) T	MCPD	7.584	8.046	602.5E6	194.9E6	75.043	73.824
7) T	MCPA	7.734	8.289	830.6E6	261.0E6	72.929	73.112
8) T	DICHLORPROP	8.109	8.656	2135.3E6	645.5E6	722.142	736.409
9) T	2,4-D	8.339	8.983	2164.6E6	627.8E6	739.588	768.045
10) T	Pentachlo...	8.637	9.507	30902.7E6	9934.1E6	766.107	754.580
11) T	2,4,5-TP ...	9.216	9.882	11900.1E6	3773.9E6	748.013	761.664
12) T	2,4,5-T	9.508	10.300	11589.5E6	3377.1E6	780.438	797.395
13) T	2,4-DB	10.083	10.864	1496.8E6	359.3E6	764.775	783.565m
14) T	DINOSEB	11.290	11.241	7940.3E6	2387.5E6	768.212m	799.714m
15) T	Picloram	11.098	12.329	14175.6E6	3635.7E6	808.042m	811.594m
16) T	DCPA	11.585	12.282	15454.9E6	4405.0E6	764.307	777.265

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030624\
Data File : PS026134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2024 21:27
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 03:09:05 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030624.M
Quant Title : 8080.M
QLast Update : Wed Mar 06 14:00:09 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

