

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030723\
 Data File : PS022333.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2023 17:53
 Operator : AR\AJ
 Sample : 01767-04MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 CE-45-030123MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/08/2023
 Supervised By : Ankita Jodhani 03/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 19:37:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:26:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.296	6.719	1049.7E6	1055.9E6	423.510m	503.053
Target Compounds							
1) T	Dalapon	2.411	2.485	1040.2E6	1711.0E6	328.108m	531.346m#
2) T	3,5-DICHL...	5.678	5.920	2013.5E6	1671.5E6	541.631	540.526
5) T	DICAMBA	6.433	6.878	4898.2E6	4603.6E6	472.043	476.371
6) T	MCPPE	6.584	6.975	368.9E6	400.5E6	48.348	52.459
7) T	MCPA	6.706	7.184	477.3E6	876.5E6	45.831	85.990 #
8) T	DICHLORPROP	7.016	7.497	1328.9E6	1336.4E6	492.081	543.608
9) T	2,4-D	7.216	7.787	1682.6E6	1669.1E6	513.197	576.414
10) T	Pentachlo...	7.439	8.206	18211.6E6	16378.3E6	505.838	505.477
11) T	2,4,5-TP ...	7.964	8.569	8316.8E6	8924.6E6	539.263	663.483
12) T	2,4,5-T	8.223	8.945	8503.5E6	6802.9E6	490.141	565.637
13) T	2,4-DB	8.730	9.449	1336.5E6	901.0E6	492.988	542.680m
14) T	DINOSEB	9.773	9.772	5122.7E6	3850.0E6	372.750	403.524
15) T	Picloram	9.610	10.742	11311.9E6	8718.3E6	459.201	444.673
16) T	DCPA	10.040	10.698	11844.3E6	8558.0E6	556.109	547.018

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030723\
Data File : PS022333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2023 17:53
Operator : AR\AJ
Sample : 01767-04MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
CE-45-030123MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/08/2023
Supervised By : Ankita Jodhani 03/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 19:37:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:26:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

