

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030821\
 Data File : PS014181.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Mar 2021 11:30
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 03:36:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030521.M
 Quant Title : 8080.M
 QLast Update : Sat Mar 06 03:17:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.900	787.8E6	297.0E6	782.326	780.680
Target Compounds							
1) T	Dalapon	3.027	2.620	811.2E6	359.3E6	691.062	681.699
2) T	3,5-DICHL...	6.421	6.057	1010.7E6	420.9E6	730.878	731.249
3) T	4-Nitroph...	7.037	6.516	360.1E6	172.4E6	693.375	702.977
5) T	DICAMBA	7.392	7.066	2981.5E6	1204.7E6	746.278	745.121
6) T	MCPD	7.560	7.160	223.2E6	99206276	75.672	74.846
7) T	MCPA	7.708	7.375	322.2E6	142.7E6	74.327	72.461
8) T	DICHLORPROP	8.082	7.691	719.6E6	299.8E6	736.981	732.682
9) T	2,4-D	8.315	7.989	797.1E6	354.1E6	737.294	733.610
10) T	Pentachlo...	8.621	8.412	11020.5E6	4094.9E6	750.960	746.103
11) T	2,4,5-TP ...	9.169	8.773	4283.6E6	1665.6E6	750.155	750.401
12) T	2,4,5-T	9.457	9.155	4463.9E6	1636.2E6	744.719	755.736
13) T	2,4-DB	10.009	9.661	765.4E6	212.7E6	723.702	742.474
14) T	DINOSEB	11.160	9.988	3418.0E6	1176.2E6	738.965	747.247
15) T	Picloram	10.996	10.984	5907.9E6	2401.9E6	740.167	727.583
16) T	DCPA	11.426	10.914	6411.5E6	2530.4E6	749.429	753.833

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030821\
Data File : PS014181.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2021 11:30
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 03:36:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030521.M
Quant Title : 8080.M
QLast Update : Sat Mar 06 03:17:26 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

