

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030921\
 Data File : PS014188.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2021 09:10
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 12:58:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030521.M
 Quant Title : 8080.M
 QLast Update : Sat Mar 06 03:17:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.901	749.4E6	281.0E6	744.184	738.530
Target Compounds							
1) T	Dalapon	3.028	2.621	788.6E6	348.1E6	671.768	660.319
2) T	3,5-DICHL...	6.421	6.058	955.6E6	396.1E6	691.027	688.029
3) T	4-Nitroph...	7.039	6.518	338.9E6	160.5E6	652.454	654.663
5) T	DICAMBA	7.392	7.066	2842.7E6	1146.3E6	711.541	708.989
6) T	MCPD	7.559	7.160	211.1E6	93428512	71.561	70.487
7) T	MCPA	7.708	7.374	304.4E6	135.4E6	70.234	68.782
8) T	DICHLORPROP	8.083	7.691	684.0E6	284.6E6	700.577	695.532
9) T	2,4-D	8.316	7.990	759.1E6	336.9E6	702.127	697.985
10) T	Pentachlo...	8.621	8.412	10597.3E6	3951.5E6	722.119	719.978
11) T	2,4,5-TP ...	9.169	8.774	4093.1E6	1591.2E6	716.791	716.886
12) T	2,4,5-T	9.457	9.155	4267.7E6	1546.6E6	711.985	714.345
13) T	2,4-DB	10.012	9.663	706.5E6	198.8E6	667.991	693.715
14) T	DINOSEB	11.161	9.988	3308.6E6	1121.8E6	715.306	712.679
15) T	Picloram	11.002	10.988	5422.6E6	2241.5E6	679.358	678.979
16) T	DCPA	11.426	10.914	6088.5E6	2416.0E6	711.673	719.763

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS030921\
Data File : PS014188.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2021 09:10
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 12:58:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030521.M
Quant Title : 8080.M
QLast Update : Sat Mar 06 03:17:26 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

