

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS031120.M
 Title : 8080.M
 Last Update : Thu Mar 12 07:13:24 2020
 Response Via : Initial Calibration

Calibration Files

750 =PS009200.D 1000 =PS009201.D 1500 =PS009202.D
 500 =PS009199.D 200 =PS009198.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	8.100	7.944	8.203	8.183	6.562	7.798 E5	8.96
2) T	3,5-DICHLOROBENZOI...	7.813	7.710	7.686	7.915	6.617	7.548 E5	7.00
3) T	4-Nitrophenol	2.477	2.468	2.510	2.481	1.999	2.387 E5	9.11
4) S	2,4-DCAA	5.733	5.702	5.834	5.714	4.676	5.532 E5	8.70
5) T	DICAMBA	2.211	2.146	2.094	2.269	1.920	2.128 E6	6.28
6) T	MCPP	2.514	2.419	2.381	2.627	2.362	2.461 E6	4.47
7) T	MCPA	3.183	3.124	3.154	3.193	2.574	3.046 E6	8.70
8) T	DICHLORPROP	5.361	5.334	5.454	5.356	4.282	5.157 E5	9.53
9) T	2,4-D	5.484	5.496	5.615	5.456	4.310	5.272 E5	10.27
10) T	Pentachlorophenol	7.226	6.753	6.177	7.775	7.224	7.031 E6	8.52
11) T	2,4,5-TP (SILVEX)	2.855	2.763	2.686	2.921	2.482	2.741 E6	6.21
12) T	2,4,5-T	2.645	2.587	2.537	2.703	2.247	2.544 E6	6.96
13) T	2,4-DB	3.610	3.681	3.861	3.520	2.626	3.460 E5	13.94
14) T	DINOSEB	2.126	2.066	2.047	2.157	1.825	2.044 E6	6.36
15) T	Picloram	3.235	3.244	3.248	3.168	2.279	3.035 E6	13.96
16) T	DCPA	3.595	3.423	3.287	3.747	3.279	3.466 E6	5.85

Signal #2 Calibration Files

750 =PS009200.D 1000 =PS009201.D 1500 =PS009202.D
 500 =PS009199.D 200 =PS009198.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	8.120	8.064	8.214	8.266	6.704	7.874 E5	8.36
2) T	3,5-DICHLOROBENZOI...	8.365	8.212	8.271	8.564	7.254	8.133 E5	6.26
3) T	4-Nitrophenol	3.285	3.255	3.300	3.327	2.743	3.182 E5	7.75
4) S	2,4-DCAA	5.626	5.593	5.707	5.692	4.843	5.492 E5	6.66
5) T	DICAMBA	2.384	2.307	2.257	2.470	2.059	2.296 E6	6.74
6) T	MCPP	1.900	1.905	1.974	1.889	1.426	1.819 E6	12.20
7) T	MCPA	2.816	2.818	2.897	2.816	2.145	2.699 E6	11.54
8) T	DICHLORPROP	5.694	5.666	5.763	5.778	4.840	5.548 E5	7.19
9) T	2,4-D	6.768	6.741	6.855	6.863	5.641	6.574 E5	7.97
10) T	Pentachlorophenol	7.792	7.286	6.605	8.468	7.798	7.590 E6	9.12
11) T	2,4,5-TP (SILVEX)	3.080	2.989	2.897	3.195	2.744	2.981 E6	5.78
12) T	2,4,5-T	3.053	2.980	2.905	3.164	2.659	2.952 E6	6.43
13) T	2,4-DB	3.897	3.927	4.072	3.926	3.159	3.796 E5	9.55
14) T	DINOSEB	2.276	2.226	2.201	2.328	1.952	2.196 E6	6.61
15) T	Picloram	4.325	4.287	4.192	4.340	3.217	4.072 E6	11.83
16) T	DCPA	3.987	3.826	3.655	4.190	3.706	3.873 E6	5.64

(#) = Out of Range