

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031120\
 Data File : PS009203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2020 22:38
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 07:15:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	779.834	814.387 E3	-4.4	101	0.00
2 T	3,5-DICHLORO BENZOIC ACID	754.798	783.717 E3	-3.8	100	0.00
3 T	4-Nitrophenol	238.713	249.656 E3	-4.6	101	0.00
4 S	2,4-DCAA	553.170	575.983 E3	-4.1	100	0.00
5 T	DICAMBA	2.128	2.218 E6	-4.2	100	0.00
6 T	MCP	2.461	2.532 E6	-2.9	101	0.00
7 T	MCPA	3.046	3.190 E6	-4.7	100	0.00
8 T	DICHLORPROP	515.743	537.798 E3	-4.3	100	0.00
9 T	2,4-D	527.225	551.465 E3	-4.6	101	0.00
10 T	Pentachlorophenol	7.031	7.209 E6	-2.5	100	0.00
11 T	2,4,5-TP (SILVEX)	2.741	2.860 E6	-4.3	100	0.00
12 T	2,4,5-T	2.544	2.659 E6	-4.5	101	0.00
13 T	2,4-DB	345.955	363.741 E3	-5.1	101	0.00
14 T	DINOSEB	2.044	2.117 E6	-3.6	100	0.00
15 T	Picloram	3.035	3.329 E6	-9.7	103	0.00
16 T	DCPA	3.466	3.598 E6	-3.8	100	0.00

Signal #2

1 T	Dalapon	787.363	812.745 E3	-3.2	100	0.00
2 T	3,5-DICHLORO BENZOIC ACID	813.307	838.916 E3	-3.1	100	0.00
3 T	4-Nitrophenol	318.198	330.718 E3	-3.9	101	0.00
4 S	2,4-DCAA	549.198	565.153 E3	-2.9	100	0.00
5 T	DICAMBA	2.296	2.384 E6	-3.8	100	0.00
6 T	MCP	1.819	1.908 E6	-4.9	100	0.00
7 T	MCPA	2.699	2.826 E6	-4.7	100	0.00
8 T	DICHLORPROP	554.822	571.033 E3	-2.9	100	0.00
9 T	2,4-D	657.361	679.499 E3	-3.4	100	0.00
10 T	Pentachlorophenol	7.590	7.736 E6	-1.9	99	0.00
11 T	2,4,5-TP (SILVEX)	2.981	3.073 E6	-3.1	100	0.00
12 T	2,4,5-T	2.952	3.055 E6	-3.5	100	0.00
13 T	2,4-DB	379.616	389.995 E3	-2.7	100	0.00
14 T	DINOSEB	2.196	2.257 E6	-2.8	99	0.00
15 T	Picloram	4.072	4.456 E6	-9.4	103	0.00
16 T	DCPA	3.873	3.974 E6	-2.6	100	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031120\
Data File : PS009203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2020 22:38
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 07:15:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0