

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031120\
 Data File : PS009201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2020 21:53
 Operator : DD\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 07:07:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:05:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.935	570.2E6	559.3E6	994.631	994.166
Target Compounds							
1) T	Dalapon	3.047	2.630	722.9E6	733.9E6	892.385	903.766
2) T	3,5-DICHL...	6.451	6.080	717.0E6	763.7E6	917.766	912.991
3) T	4-Nitroph...	7.071	6.536	224.6E6	296.2E6	906.853	901.525
5) T	DICAMBA	7.435	7.105	2017.7E6	2168.7E6	912.621	909.542
6) T	MCPD	7.608	7.206	227.4E6	179.0E6	90.436	94.230
7) T	MCPA	7.758	7.423	290.5E6	262.1E6	91.251	93.059
8) T	DICHLORPROP	8.130	7.733	501.4E6	532.6E6	935.334	935.491
9) T	2,4-D	8.358	8.027	516.6E6	633.7E6	942.002	936.330
10) T	Pentachlo...	8.671	8.461	6415.8E6	6922.1E6	887.860	888.393
11) T	2,4,5-TP ...	9.222	8.822	2624.4E6	2839.4E6	919.231	921.783
12) T	2,4,5-T	9.505	9.200	2457.6E6	2830.6E6	929.307	927.036
13) T	2,4-DB	10.059	9.708	349.6E6	373.0E6	968.589	957.237
14) T	DINOSEB	11.225	10.044	1942.1E6	2092.6E6	913.590	919.403
15) T	Picloram	11.037	11.023	3082.0E6	4072.6E6	952.648	941.617
16) T	DCPA	11.489	10.974	3285.7E6	3673.0E6	913.910	921.252

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031120\
Data File : PS009201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2020 21:53
Operator : DD\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 07:07:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:05:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

