

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031219\
 Data File : PS003508.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2019 09:19
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/13/2019 10:09:25 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 04:37:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.279	10.778	336.1E6	348.4E6	765.586	741.967
Target Compounds							
1) T	Dalapon	2.841	2.313	452.7E6	424.3E6	696.476	670.092
2) T	3,5-DICHL...	9.706	8.939	484.2E6	427.6E6	726.375	692.542
3) T	4-Nitroph...	10.880	9.933	150.0E6	150.2E6	712.835	668.242
5) T	DICAMBA	11.580	11.057	1279.8E6	1182.8E6	726.017	707.965
6) T	MCPD	11.922	11.288	97307228	85430588	74.116m	72.373m
7) T	MCPA	12.171	11.664	144.4E6	123.0E6	72.802m	71.308
8) T	DICHLORPROP	12.773	12.183	312.2E6	289.7E6	721.493	701.046
9) T	2,4-D	13.137	12.657	296.4E6	363.4E6	639.086m	692.315
10) T	Pentachlo...	13.495	13.186	5137.8E6	4697.3E6	727.287m	713.250m
11) T	2,4,5-TP ...	14.375	13.814	1979.3E6	1833.2E6	736.807	719.901
12) T	2,4,5-T	14.779	14.379	1913.5E6	1439.9E6	724.342	713.964
13) T	2,4-DB	15.528	15.031	166.5E6	183.6E6	703.312	719.008
14) T	DINOSEB	16.985	15.452	1641.1E6	1273.4E6	730.942	713.440
15) T	Picloram	16.771	16.688	1803.7E6	1753.8E6	732.843	690.977
16) T	DCPA	17.331	16.627	2409.4E6	2279.7E6	731.194m	732.943

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031219\
Data File : PS003508.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2019 09:19
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
3/13/2019 10:09:25 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 04:37:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

