

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031220\
 Data File : PS009250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2020 05:22
 Operator : DD\AJ
 Sample : L1864-05
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 07:08:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.934	284.2E6	255.0E6	513.743	464.404
Target Compounds							
1) T	Dalapon	3.010	2.590	18214174	92008238	23.356	116.856 #
2) T	3,5-DICHL...	0.000	6.110	0	81616254	N.D.	100.351
3) T	4-Nitroph...	7.010	6.592	23118563	15149726	96.847	47.611 #
8) T	DICHLORPROP	8.117	7.753	-2003215	39723324	N.D.	71.597
9) T	2,4-D	8.295	0.000	13323743	0	25.271	N.D. #
12) T	2,4,5-T	9.444	0.000	27385922	0	10.766	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031220\
Data File : PS009250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2020 05:22
Operator : DD\AJ
Sample : L1864-05
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 07:08:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.010 min
Delta R.T.: -0.036 min
Response: 18214174
Conc: 23.36 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.590 min
Delta R.T.: -0.037 min
Response: 92008238
Conc: 116.86 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.451 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.110 min
Delta R.T.: 0.030 min
Response: 81616254
Conc: 100.35 ng/ml



#3 4-Nitrophenol


R.T.: 7.010 min
Delta R.T.: -0.061 min
Response: 23118563
Conc: 96.85 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.592 min
Delta R.T.: 0.056 min
Response: 15149726
Conc: 47.61 ng/ml




#4 2,4-DCAA

R.T.: 7.242 min
Delta R.T.: 0.000 min
Response: 284187380
Conc: 513.74 ng/ml



#4 2,4-DCAA


R.T.: 6.934 min
Delta R.T.: -0.001 min
Response: 255049487
Conc: 464.40 ng/ml



#8 DICHLORPROP


R.T.: 8.117 min
Delta R.T.: -0.012 min
Response: -2003215
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.753 min
Delta R.T.: 0.020 min
Response: 39723324
Conc: 71.60 ng/ml




#9 2,4-D

R.T.: 8.295 min
Delta R.T.: -0.062 min
Response: 13323743
Conc: 25.27 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.027 min
Response: 0
Conc: N.D.



#12 2,4,5-T

R.T.: 9.444 min
Delta R.T.: -0.061 min
Response: 27385922
Conc: 10.77 ng/ml

Instrument :
ECD_S
ClientSampleId :



#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.201 min
Response: 0
Conc: N.D.