

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031220\
 Data File : PS009263.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2020 10:21
 Operator : DD\AJ
 Sample : L1887-03
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 11:06:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.934	301.1E6	318.3E6	544.228	579.606
Target Compounds							
1) T	Dalapon	3.008	2.591	17843622	88208456	22.881	112.030 #
2) T	3,5-DICHL...	0.000	6.110	0	62094985	N.D.	76.349
3) T	4-Nitroph...	7.010	6.592	15304411	9971109	64.112	31.336 #
8) T	DICHLORPROP	8.118	7.754	-2213739	31117357	N.D.	56.085
9) T	2,4-D	8.427	0.000	9438388	0	17.902	N.D. #
12) T	2,4,5-T	9.443	0.000	14941664	0	5.874	N.D. #
15) T	Picloram	0.000	11.022	0	32512699	N.D.	7.984
16) T	DCPA	0.000	11.022	0	32512699	N.D.	8.395



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031220\
Data File : PS009263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2020 10:21
Operator : DD\AJ
Sample : L1887-03
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 11:06:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.008 min
Delta R.T.: -0.038 min
Response: 17843622
Conc: 22.88 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.591 min
Delta R.T.: -0.035 min
Response: 88208456
Conc: 112.03 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.451 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.110 min
Delta R.T.: 0.031 min
Response: 62094985
Conc: 76.35 ng/ml



#3 4-Nitrophenol


R.T.: 7.010 min
Delta R.T.: -0.061 min
Response: 15304411
Conc: 64.11 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.592 min
Delta R.T.: 0.055 min
Response: 9971109
Conc: 31.34 ng/ml




#4 2,4-DCAA

R.T.: 7.242 min
Delta R.T.: 0.000 min
Response: 301050602
Conc: 544.23 ng/ml



#4 2,4-DCAA


R.T.: 6.934 min
Delta R.T.: -0.001 min
Response: 318318437
Conc: 579.61 ng/ml



#8 DICHLORPROP


R.T.: 8.118 min
Delta R.T.: -0.011 min
Response: -2213739
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.754 min
Delta R.T.: 0.021 min
Response: 31117357
Conc: 56.09 ng/ml




#9 2,4-D

R.T.: 8.427 min
Delta R.T.: 0.069 min
Response: 9438388
Conc: 17.90 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.027 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 9.443 min
Delta R.T.: -0.062 min
Response: 14941664
Conc: 5.87 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.201 min
Response: 0
Conc: N.D.




#15 Picloram

R.T.: 0.000 min
Exp R.T. : 11.037 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 11.022 min
Delta R.T.: 0.000 min
Response: 32512699
Conc: 7.98 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.489 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 11.022 min
Delta R.T.: 0.048 min
Response: 32512699
Conc: 8.39 ng/ml