

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031419\
 Data File : PS003564.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2019 18:03
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/15/2019 3:13:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 01:11:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.272	10.775	331.8E6	348.0E6	755.743	741.153
Target Compounds							
1) T	Dalapon	2.841	2.315	443.9E6	425.7E6	682.933	672.319
2) T	3,5-DICHL...	9.701	8.936	469.7E6	425.4E6	704.717	688.900
3) T	4-Nitroph...	10.866	9.928	131.8E6	151.4E6	626.620m	673.282
5) T	DICAMBA	11.576	11.054	1277.8E6	1181.7E6	724.896	707.314
6) T	MCPD	11.918	11.287	97088303	85286619	73.949m	72.252m
7) T	MCPA	12.168	11.662	143.1E6	123.2E6	72.132m	71.415
8) T	DICHLORPROP	12.767	12.180	311.5E6	289.0E6	720.061	699.308
9) T	2,4-D	13.129	12.653	354.0E6	380.1E6	763.309m	724.105
10) T	Pentachlo...	13.492	13.184	5168.7E6	4727.7E6	731.650	717.880
11) T	2,4,5-TP ...	14.369	13.811	1991.0E6	1851.2E6	741.144	726.959
12) T	2,4,5-T	14.770	14.375	1971.0E6	1494.1E6	746.122	740.857
13) T	2,4-DB	15.517	15.027	181.2E6	204.6E6	765.225m	801.068
14) T	DINOSEB	16.981	15.449	1583.7E6	1282.2E6	705.362	718.377
15) T	Picloram	16.759	16.682	2314.1E6	2209.5E6	940.237	870.509
16) T	DCPA	17.328	16.623	2432.0E6	2314.1E6	738.038	744.004

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031419\
Data File : PS003564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2019 18:03
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
Client Sampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
3/15/2019 3:13:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 01:11:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

