

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031422\
 Data File : PS018556.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2022 10:06
 Operator : AJ\MA
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 11:37:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:36:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.812	1442.4E6	353.6E6	522.149	516.696
Target Compounds							
1) T	Dalapon	3.031	2.533	1587.4E6	376.4E6	457.925	468.888
2) T	3,5-DICHL...	6.443	5.984	1825.7E6	496.7E6	487.019	486.175
3) T	4-Nitroph...	7.061	6.422	780.6E6	188.5E6	465.110	479.167
5) T	DICAMBA	7.428	6.980	5285.9E6	1450.3E6	492.813	493.010
6) T	MCPD	7.599	7.077	393.3E6	114.7E6	46.857	47.975
7) T	MCPA	7.748	7.290	560.9E6	162.8E6	47.228	47.983
8) T	DICHLORPROP	8.124	7.603	1312.1E6	374.9E6	491.005	487.893
9) T	2,4-D	8.351	7.894	1602.6E6	444.9E6	490.060	486.664
10) T	Pentachlo...	8.664	8.326	18700.4E6	5034.5E6	519.537	524.846
11) T	2,4,5-TP ...	9.217	8.686	7516.5E6	2163.0E6	504.585	503.810
12) T	2,4,5-T	9.500	9.062	7380.7E6	2047.8E6	503.893	502.036
13) T	2,4-DB	10.052	9.569	1306.7E6	269.5E6	490.261	486.581
14) T	DINOSEB	11.219	9.899	5212.4E6	1477.9E6	497.079	489.938
15) T	Picloram	11.030	10.875	10224.3E6	2640.1E6	501.262	501.851
16) T	DCPA	11.486	10.832	9114.3E6	2648.7E6	519.971	514.963

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031422\
Data File : PS018556.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2022 10:06
Operator : AJ\MA
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 11:37:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:36:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

