

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031422\
 Data File : PS018576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2022 18:59
 Operator : AJ\MA
 Sample : N1645-07
 Misc : LOD 150 PPB
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 01:06:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.811	910.3E6	233.3E6	319.868	331.611
Target Compounds							
1) T	Dalapon	3.031	2.535	595.1E6	113.3E6	153.043	137.138
2) T	3,5-DICHL...	6.441	5.984	591.0E6	146.1E6	153.358	139.690
3) T	4-Nitroph...	7.061	6.421	234.5E6	62045764	136.077	152.719
5) T	DICAMBA	7.427	6.979	1615.9E6	417.4E6	148.156	140.121
6) T	MCPD	7.594	7.072	104.9E6	31328627	12.357	12.797
7) T	MCPA	7.741	7.283	158.6E6	47882524	13.111	13.712
8) T	DICHLORPROP	8.123	7.603	408.1E6	128.4E6	148.156	162.219
9) T	2,4-D	8.350	7.893	488.3E6	133.2E6	145.721	142.259
10) T	Pentachlo...	8.662	8.324	5998.3E6	1621.4E6	165.688	165.103
11) T	2,4,5-TP ...	9.216	8.685	2323.9E6	650.8E6	152.807	149.203
12) T	2,4,5-T	9.498	9.061	2267.1E6	619.1E6	151.763	149.473
13) T	2,4-DB	10.051	9.569	376.8E6	79871228	137.748	140.681
14) T	DINOSEB	11.217	9.898	1536.6E6	438.4E6	143.701	143.079
15) T	Picloram	11.029	10.874	3013.7E6	803.7E6	146.609	151.173
16) T	DCPA	11.484	10.831	2746.1E6	821.1E6	153.494	156.128

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031422\
Data File : PS018576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2022 18:59
Operator : AJ\MA
Sample : N1645-07
Misc : LOD 150 PPB
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
LOD-MDL-WATER-01-QT1-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 01:06:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

