

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031422\
 Data File : PS018585.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2022 23:45
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 01:10:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.812	2009.8E6	509.2E6	706.190	723.629
Target Compounds							
1) T	Dalapon	3.032	2.534	2893.4E6	535.4E6	744.087	647.875
2) T	3,5-DICHL...	6.442	5.984	2608.3E6	694.3E6	676.862	663.674
3) T	4-Nitroph...	7.060	6.422	1104.5E6	265.9E6	640.964	654.579
5) T	DICAMBA	7.427	6.980	7252.2E6	2024.7E6	664.931	679.760
6) T	MCPD	7.600	7.079	560.6E6	159.9E6	66.037	65.298
7) T	MCPA	7.749	7.293	810.6E6	232.5E6	67.020	66.581
8) T	DICHLORPROP	8.122	7.603	1812.4E6	604.5E6	658.004	763.676
9) T	2,4-D	8.350	7.894	2189.3E6	641.4E6	653.394	685.231
10) T	Pentachlo...	8.663	8.325	24334.0E6	6725.5E6	672.162	684.859
11) T	2,4,5-TP ...	9.216	8.686	10069.1E6	3044.0E6	662.086	697.853
12) T	2,4,5-T	9.499	9.062	9942.3E6	2925.1E6	665.541	706.175
13) T	2,4-DB	10.052	9.569	1865.9E6	398.5E6	682.116	701.943
14) T	DINOSEB	11.217	9.898	6949.0E6	2131.7E6	649.853	695.768
15) T	Picloram	11.030	10.875	13883.8E6	3867.2E6	675.421	727.449
16) T	DCPA	11.484	10.832	11848.8E6	3708.0E6	662.287	705.036

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031422\
Data File : PS018585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2022 23:45
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 01:10:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

