

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL031521\  
 Data File : PL065296.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 15 Mar 2021 09:11  
 Operator : AR\AJ  
 Sample : HEXANE  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 16 02:03:50 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL031321.M  
 Quant Title : GC Extractables  
 QLast Update : Sat Mar 13 01:11:50 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

| Compound | RT#1 | RT#2 | Resp#1 | Resp#2 | ng/ml | ng/ml |
|----------|------|------|--------|--------|-------|-------|
|----------|------|------|--------|--------|-------|-------|

#### System Monitoring Compounds

#### Target Compounds

|      |          |       |       |            |      |         |
|------|----------|-------|-------|------------|------|---------|
| 6) B | beta-BHC | 0.000 | 4.748 | 0 12203038 | N.D. | 6.992 # |
|------|----------|-------|-------|------------|------|---------|


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL031521\  
Data File : PL065296.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 15 Mar 2021 09:11  
Operator : AR\AJ  
Sample : HEXANE  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 16 02:03:50 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL031321.M  
Quant Title : GC Extractables  
QLast Update : Sat Mar 13 01:11:50 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min  
Exp R.T. : 3.295 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_S  
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min  
Exp R.T. : 3.883 min  
Response: 0  
Conc: N.D.




#6 beta-BHC

R.T.: 0.000 min  
Exp R.T. : 4.241 min  
Response: 0  
Conc: N.D.



#6 beta-BHC


R.T.: 4.748 min  
Delta R.T.: 0.007 min  
Response: 12203038  
Conc: 6.99 ng/ml



#28 Decachlorobiphenyl

R.T.: 0.000 min  
Exp R.T. : 7.941 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_S  
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 0.000 min  
Exp R.T. : 8.892 min  
Response: 0  
Conc: N.D.