

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031522\
 Data File : PS018611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2022 20:00
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 21:46:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.232	6.811	2152.6E6	523.2E6	756.386	743.482
Target Compounds							
1) T	Dalapon	3.030	2.533	2622.6E6	540.9E6	674.462	654.508
2) T	3,5-DICHL...	6.441	5.983	2782.4E6	718.2E6	722.038	686.516
3) T	4-Nitroph...	7.059	6.421	1176.7E6	270.7E6	682.915	666.337
5) T	DICAMBA	7.425	6.979	7781.2E6	2084.6E6	713.434	699.883
6) T	MCP	7.598	7.078	603.5E6	165.3E6	71.100	67.526
7) T	MCPA	7.748	7.291	848.4E6	239.3E6	70.149	68.534
8) T	DICHLORPROP	8.121	7.602	1945.7E6	571.1E6	706.408	721.484
9) T	2,4-D	8.349	7.892	2370.2E6	659.0E6	707.382	704.076
10) T	Pentachlo...	8.661	8.324	26088.3E6	6934.7E6	720.619	706.162
11) T	2,4,5-TP ...	9.213	8.684	10897.0E6	3124.3E6	716.520	716.245
12) T	2,4,5-T	9.497	9.061	10827.4E6	3004.3E6	724.786	725.303
13) T	2,4-DB	10.049	9.568	2096.4E6	410.1E6	766.376	722.315
14) T	DINOSEB	11.215	9.897	7845.4E6	2186.3E6	733.685	713.594
15) T	Picloram	11.028	10.874	15099.6E6	3834.6E6	734.570	721.325
16) T	DCPA	11.482	10.831	13361.6E6	3801.6E6	746.843	722.846

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031522\
Data File : PS018611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2022 20:00
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 21:46:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

