

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031524\
 Data File : PS026161.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2024 09:53
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/18/2024
 Supervised By : Ankita Jodhani 03/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 04:22:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030624.M
 Quant Title : 8080.M
 QLast Update : Wed Mar 06 14:00:09 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.200	7.724	1978.7E6	558.9E6	731.480	720.903
Target Compounds							
1) T	Dalapon	2.645	2.693	2751.9E6	993.8E6	636.789	620.498
2) T	3,5-DICHL...	6.377	6.686	2614.9E6	787.4E6	671.303m	666.129
3) T	4-Nitroph...	6.999	7.253	1004.3E6	256.1E6	747.250	706.525
5) T	DICAMBA	7.386	7.921	7899.2E6	2416.6E6	695.588	684.360
6) T	MCPD	7.568	8.024	564.8E6	181.8E6	70.349	68.833
7) T	MCPA	7.717	8.266	785.6E6	242.4E6	68.980	67.913
8) T	DICHLORPROP	8.092	8.633	2035.4E6	606.5E6	688.360	691.946
9) T	2,4-D	8.321	8.959	2134.3E6	582.7E6	729.228	712.781
10) T	Pentachlo...	8.618	9.481	29699.6E6	9184.5E6	736.280	697.637
11) T	2,4,5-TP ...	9.196	9.856	11440.2E6	3496.7E6	719.106	705.726
12) T	2,4,5-T	9.488	10.271	11263.6E6	3153.5E6	758.494	744.589m
13) T	2,4-DB	10.061	10.836	1462.9E6	340.5E6	747.441	742.447
14) T	DINOSEB	11.267	11.212	8182.0E6	2340.2E6	791.597	783.862
15) T	Picloram	11.075	12.295	12936.3E6	3183.7E6	737.400m	710.687m
16) T	DCPA	11.562	12.250	14709.7E6	4118.1E6	727.454m	726.640m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031524\
Data File : PS026161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2024 09:53
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/18/2024
Supervised By : Ankita Jodhani 03/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 04:22:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS030624.M
Quant Title : 8080.M
QLast Update : Wed Mar 06 14:00:09 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

