

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031721\
 Data File : PS014273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Mar 2021 11:23
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS031721

Manual Integrations
 APPROVED

Ankita
 3/18/2021 3:59:44 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 02:24:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.167	6.888	1107.3E6	365.9E6	774.982	775.505m
Target Compounds							
1) T	Dalapon	3.006	2.599	1115.1E6	417.8E6	716.405	705.564
2) T	3,5-DICHL...	6.391	6.049	1392.8E6	542.7E6	721.478	728.887
3) T	4-Nitroph...	7.001	6.503	540.4E6	227.4E6	723.641	748.471
5) T	DICAMBA	7.356	7.055	3956.1E6	1475.6E6	735.634	736.572
6) T	MCPD	7.523	7.151	307.8E6	124.0E6	74.024	74.448
7) T	MCPA	7.671	7.366	429.4E6	180.7E6	74.631	74.177
8) T	DICHLORPROP	8.041	7.681	972.0E6	374.7E6	734.024	735.463
9) T	2,4-D	8.270	7.977	1099.4E6	417.0E6	745.885	698.396
10) T	Pentachlo...	8.577	8.401	14789.3E6	5026.6E6	743.011	736.002
11) T	2,4,5-TP ...	9.122	8.762	5754.1E6	2034.3E6	746.752	747.734
12) T	2,4,5-T	9.407	9.142	5862.5E6	2014.7E6	753.443	755.685
13) T	2,4-DB	9.954	9.647	1190.0E6	260.3E6	769.556	734.018
14) T	DINOSEB	11.109	9.977	4323.5E6	1383.6E6	756.432	765.663
15) T	Picloram	10.932	10.960	8276.4E6	3099.0E6	792.579	816.892
16) T	DCPA	11.375	10.903	8887.8E6	3091.1E6	754.596	753.580

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031721\
Data File : PS014273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2021 11:23
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS031721

Manual Integrations
APPROVED

Ankita
3/18/2021 3:59:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 02:24:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

