

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031723\
 Data File : PS022426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2023 07:45
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/20/2023
 Supervised By : Ankita Jodhani 03/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 09:11:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
 Quant Title : 8080.M
 QLast Update : Mon Feb 20 13:26:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.298	6.720	1910.6E6	1673.7E6	770.801	797.414
Target Compounds							
1) T	Dalapon	2.410	2.483	1962.2E6	2082.8E6	618.949	646.788
2) T	3,5-DICHL...	5.679	5.921	2696.3E6	2400.0E6	725.313	776.106
3) T	4-Nitroph...	6.142	6.348	1024.2E6	1048.0E6	804.887	784.686
5) T	DICAMBA	6.434	6.878	7673.3E6	7567.8E6	739.481	783.097
6) T	MCPD	6.587	6.978	586.6E6	614.5E6	76.874	80.491
7) T	MCPA	6.709	7.188	786.3E6	804.2E6	75.505	78.898
8) T	DICHLORPROP	7.017	7.499	1969.2E6	1912.7E6	729.179	778.025
9) T	2,4-D	7.219	7.789	2477.5E6	2258.7E6	755.620	780.013
10) T	Pentachlo...	7.439	8.207	26775.6E6	25293.3E6	743.706	780.615
11) T	2,4,5-TP ...	7.966	8.571	11359.8E6	10599.8E6	736.574	788.028
12) T	2,4,5-T	8.225	8.945	12407.1E6	9976.1E6	715.142	829.469
13) T	2,4-DB	8.730	9.450	2159.9E6	1317.0E6	796.751	793.265m
14) T	DINOSEB	9.773	9.773	9735.6E6	7504.1E6	708.409	786.512
15) T	Picloram	9.612	10.743	18003.1E6	14760.5E6	730.825	752.854
16) T	DCPA	10.040	10.698	15619.9E6	12465.6E6	733.379	796.785

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031723\
Data File : PS022426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2023 07:45
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/20/2023
Supervised By : Ankita Jodhani 03/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 09:11:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022023.M
Quant Title : 8080.M
QLast Update : Mon Feb 20 13:26:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

