

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031725\
 Data File : PS029443.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Mar 2025 10:56
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/18/2025
 Supervised By : mohammad ahmed 03/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 01:42:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022125.M
 Quant Title : 8080.M
 QLast Update : Sat Feb 22 01:02:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.155	7.617	3130.5E6	647.4E6	759.722	688.512
Target Compounds							
1) T	Dalapon	2.602	2.644	4017.5E6	1447.1E6	727.775	683.530
2) T	3,5-DICHL...	6.338	6.593	4170.6E6	922.1E6	712.780	651.125
3) T	4-Nitroph...	6.958	7.150	1437.7E6	488.9E6	614.344	612.966
5) T	DICAMBA	7.337	7.809	12848.7E6	3542.1E6	739.879	687.398m
6) T	MCP	7.518	7.915	861.7E6	151.7E6	69.913	64.852
7) T	MCPA	7.665	8.153	1204.5E6	208.6E6	71.083	65.412
8) T	DICHLORPROP	8.037	8.518	3275.3E6	855.5E6	712.561	669.007
9) T	2,4-D	8.268	8.843	3267.2E6	814.6E6	687.779	641.448
10) T	Pentachlo...	8.562	9.358	44219.9E6	16567.1E6	757.759	719.912
11) T	2,4,5-TP ...	9.133	9.736	18153.2E6	6402.1E6	738.329	708.452
12) T	2,4,5-T	9.425	10.150	18328.3E6	5521.7E6	719.074	672.537
13) T	2,4-DB	9.994	10.712	2589.3E6	509.3E6	620.121	665.244m
14) T	DINOSEB	11.185	11.087	11817.4E6	4043.0E6	710.573m	667.000m
15) T	Picloram	11.002	12.161	18644.0E6	6776.8E6	610.407	586.102
16) T	DCPA	11.482	12.120	22101.4E6	8586.3E6	736.627	725.410m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031725\
Data File : PS029443.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2025 10:56
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/18/2025
Supervised By : mohammad ahmed 03/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 01:42:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022125.M
Quant Title : 8080.M
QLast Update : Sat Feb 22 01:02:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

