

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031920\
 Data File : PS009364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2020 17:04
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 18:11:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.241	6.933	489.3E6	440.7E6	884.554	802.374
Target Compounds							
1) T	Dalapon	3.048	2.630	558.0E6	493.9E6	715.572	627.327
2) T	3,5-DICHL...	6.451	6.078	621.6E6	605.1E6	823.573	743.979
3) T	4-Nitroph...	7.070	6.535	193.6E6	230.5E6	810.996	724.356
5) T	DICAMBA	7.434	7.103	1721.3E6	1681.8E6	808.814	732.633
6) T	MCP	7.606	7.202	138.3E6	144.8E6	56.191	79.612 #
7) T	MCPA	7.755	7.419	193.3E6	197.7E6	63.471	73.248
8) T	DICHLORPROP	8.128	7.730	416.3E6	420.7E6	807.107	758.194
9) T	2,4-D	8.357	8.024	423.5E6	497.2E6	803.226	756.419
10) T	Pentachlo...	8.671	8.457	5513.8E6	5350.7E6	784.190	704.965
11) T	2,4,5-TP ...	9.219	8.818	2227.6E6	2251.8E6	812.631	755.360
12) T	2,4,5-T	9.503	9.197	2047.7E6	2224.4E6	804.967	753.468
13) T	2,4-DB	10.057	9.705	273.4E6	294.7E6	790.392	776.207
14) T	DINOSEB	11.222	10.040	1644.7E6	1665.1E6	804.574	758.055
15) T	Picloram	11.036	11.019	2182.5E6	2777.0E6	719.131	681.937
16) T	DCPA	11.487	10.970	2757.0E6	2927.9E6	795.382	755.961

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031920\
Data File : PS009364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2020 17:04
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 18:11:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

